



29 July 2020

C0G4569

Paul Hughes

Eastex Environmental Laboratory - Nacogdoches


1119 South University Drive

Nacogdoches, TX 75961

RE: Lily Grove WQP

Water Quality Parameters (WQP) Report

Enclosed, is your Water Quality Parameter (WQP) report with the necessary support data. Eastex Lab has reported your results to TCEQ in the following formats: 1. EDD of your data and 2. PDF of your report.



Here are your next actions:

1. Check your report for any errors, such as: Regulatory ID, site names. Please make sure all your sites are listed correctly.
2. Evaluate your WQP data for corrosivity and implement your plan.
3. Please retain all forms and reports from the lab for future field inspections.

If you have questions regarding any of the enclosed data, please call me at 936-653-3249. For compliance questions, please call TCEQ's WQP team at 512-239-4691.

Daniel Bowen
Operations Manager



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Eastex Environmental Laboratory - Nacogdoches
1119 South University Drive
Nacogdoches TX, 75961

TCEQ Report ID: **PWS_TX1740014_AC_20200714_WQP Analysis Report**

Project: **Lily Grove WQP**
Laboratory Sample ID: **C0G4569-01**
Sampling Point: **EWQP**
Sample Location: **N0G0813-01 (Martin Plant)**
State Assigned ID: **PBCU005**

Sample Received Date and Time: 07/21/2020 12:42
Sample Date and Time: 07/14/2020 10:40
Sample Matrix: Drinking
Sample Type: Grab
Collector: Boyd Dueboay

Analyte	Result	Units	Reporting Limits	MDL	Nelac Status	Batch	Analyzed	Analyst	Method	Notes
Eastex Environmental Laboratory - Coldspring										
Calcium, Total	9.47	mg/L	2.50	0.0879	P	B0G3559	07/28/2020 16:36	LAN	EPA 200.7	12
Chloride	16.9	mg/L	5.0	1.9	A	B0G3445	07/24/2020 15:00	AMS	EPA 300.0	
Iron, Total	0.390	mg/L	0.150	0.00255	A	B0G3557	07/28/2020 12:02	LAN	EPA 200.7	
Manganese, Total	<0.0100	mg/L	0.0100	0.00310	A	B0G3557	07/28/2020 12:02	LAN	EPA 200.7	
Sodium, Total	85.6	mg/L	2.50	0.111	A	B0G3559	07/28/2020 16:36	LAN	EPA 200.7	
Sulfate	49.6	mg/L	4.0	2.0	A	B0G3445	07/24/2020 15:00	AMS	EPA 300.0	



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Eastex Environmental Laboratory - Nacogdoches
1119 South University Drive
Nacogdoches TX, 75961

TCEQ Report ID: **PWS_TX1740014_AC_20200714_WQP Analysis Report**

Project: **Lily Grove WQP**
Laboratory Sample ID: **C0G4569-02**
Sampling Point: **EWQP**
Sample Location: **N0G0813-02 (Nat Plant)**
State Assigned ID: **PBCU002**

Sample Received Date and Time: 07/21/2020 12:42
Sample Date and Time: 07/14/2020 10:25
Sample Matrix: Drinking
Sample Type: Grab
Collector: Boyd Dueboay

Analyte	Result	Units	Reporting Limits	MDL	Nelac Status	Batch	Analyzed	Analyst	Method	Notes
---------	--------	-------	------------------	-----	--------------	-------	----------	---------	--------	-------

Eastex Environmental Laboratory - Coldspring

Calcium, Total	8.46	mg/L	2.50	0.0879	P	B0G3559	07/28/2020 16:38	LAN	EPA 200.7	12
Chloride	14.4	mg/L	5.0	1.9	A	B0G3445	07/24/2020 15:00	AMS	EPA 300.0	
Iron, Total	<0.150	mg/L	0.150	0.00255	A	B0G3557	07/28/2020 12:07	LAN	EPA 200.7	
Manganese, Total	<0.0100	mg/L	0.0100	0.00310	A	B0G3557	07/28/2020 12:07	LAN	EPA 200.7	
Sodium, Total	15.7	mg/L	2.50	0.111	A	B0G3559	07/28/2020 16:38	LAN	EPA 200.7	
Sulfate	34.6	mg/L	4.0	2.0	A	B0G3445	07/24/2020 15:00	AMS	EPA 300.0	



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Eastex Environmental Laboratory - Nacogdoches
1119 South University Drive
Nacogdoches TX, 75961

TCEQ Report ID: **PWS_TX1740014_AC_20200714_WQP Analysis Report**

Project: **Lily Grove WQP**
Laboratory Sample ID: **C0G4569-03**
Sampling Point: **DSTWQP**
Sample Location: **N0G0813-03 (366 CR 820)**
State Assigned ID: **DS01**

Sample Received Date and Time: 07/21/2020 12:42
Sample Date and Time: 07/14/2020 11:10
Sample Matrix: Drinking
Sample Type: Grab
Collector: Boyd Dueboay

Analyte	Result	Units	Reporting Limits	MDL	Nelac Status	Batch	Analyzed	Analyst	Method	Notes
Eastex Environmental Laboratory - Coldspring										
Calcium, Total	10.9	mg/L	2.50	0.0879	P	B0G3559	07/28/2020 16:40	LAN	EPA 200.7	12
Chloride	14.3	mg/L	5.0	1.9	A	B0G3445	07/24/2020 15:00	AMS	EPA 300.0	
Iron, Total	0.916	mg/L	0.150	0.00255	A	B0G3557	07/28/2020 12:12	LAN	EPA 200.7	
Manganese, Total	0.0119	mg/L	0.0100	0.00310	A	B0G3557	07/28/2020 12:12	LAN	EPA 200.7	
Sodium, Total	76.9	mg/L	2.50	0.111	A	B0G3559	07/28/2020 16:40	LAN	EPA 200.7	
Sulfate	34.6	mg/L	4.0	2.0	A	B0G3445	07/24/2020 15:00	AMS	EPA 300.0	



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Eastex Environmental Laboratory - Nacogdoches
1119 South University Drive
Nacogdoches TX, 75961

TCEQ Report ID: **PWS_TX1740014_AC_20200714_WQP Analysis Report**

Project: **Lily Grove WQP**
Laboratory Sample ID: **C0G4569-04**
Sampling Point: **DSTWQP**
Sample Location: **N0G0813-04 (12232 FM 343)**
State Assigned ID: **DS01**

Sample Received Date and Time: 07/21/2020 12:42
Sample Date and Time: 07/14/2020 10:35
Sample Matrix: Drinking
Sample Type: Grab
Collector: Boyd Dueboay

Analyte	Result	Units	Reporting Limits	MDL	Nelac Status	Batch	Analyzed	Analyst	Method	Notes
Eastex Environmental Laboratory - Coldspring										
Calcium, Total	7.99	mg/L	2.50	0.0879	P	B0G3559	07/28/2020 16:42	LAN	EPA 200.7	12
Chloride	32.2	mg/L	5.0	1.9	A	B0G3445	07/24/2020 15:00	AMS	EPA 300.0	
Iron, Total	<0.150	mg/L	0.150	0.00255	A	B0G3557	07/28/2020 12:16	LAN	EPA 200.7	
Manganese, Total	<0.0100	mg/L	0.0100	0.00310	A	B0G3557	07/28/2020 12:16	LAN	EPA 200.7	
Sodium, Total	14.5	mg/L	2.50	0.111	A	B0G3559	07/28/2020 16:42	LAN	EPA 200.7	
Sulfate	4.8	mg/L	4.0	2.0	A	B0G3445	07/24/2020 15:00	AMS	EPA 300.0	



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Eastex Environmental Laboratory - Nacogdoches
1119 South University Drive
Nacogdoches TX, 75961

TCEQ Report ID: **PWS_TX1740014_AC_20200714_WQP Analysis Report**

EPA 300.0 - Quality Control

Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch B0G3445 - No Prep

Blank (B0G3445-BLK1)

Prepared & Analyzed: 07/24/20

Chloride	ND	5.0	mg/L
Sulfate	ND	4.0	mg/L

LCS (B0G3445-BS1)

Prepared & Analyzed: 07/24/20

Chloride	25.8		mg/L	25.0		103	90-110
Sulfate	20.9		mg/L	20.0		104	90-110

MRL Check (B0G3445-MRL1)

Prepared & Analyzed: 07/24/20

Chloride	5.51		mg/L	5.00		110	70-130
Sulfate	4.40		mg/L	4.00		110	70-130

Matrix Spike (B0G3445-MS1)

Source: C0G3802-02

Prepared & Analyzed: 07/24/20

Chloride	157	5.0	mg/L	125	40.1	93.2	80-120
Sulfate	96.8	4.0	mg/L	100	ND	96.8	80-120

Matrix Spike Dup (B0G3445-MSD1)

Source: C0G3802-02

Prepared & Analyzed: 07/24/20

Chloride	157	5.0	mg/L	125	40.1	93.3	80-120	0.0668	20
Sulfate	96.9	4.0	mg/L	100	ND	96.9	80-120	0.197	20

Batch B0G3557 - EPA 200.7

Blank (B0G3557-BLK1)

Prepared: 07/27/20 Analyzed: 07/28/20

Iron, Total	ND	0.150	mg/L
Manganese, Total	ND	0.0100	mg/L

Blank (B0G3557-BLK2)

Prepared: 07/27/20 Analyzed: 07/28/20

Iron, Total	ND	0.150	mg/L
Manganese, Total	ND	0.0100	mg/L



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Eastex Environmental Laboratory - Nacogdoches
1119 South University Drive
Nacogdoches TX, 75961

TCEQ Report ID: **PWS_TX1740014_AC_20200714_WQP Analysis Report**

EPA 200.7 - Quality Control

Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B0G3557 - EPA 200.7										
Blank (B0G3557-BLK3)										
Prepared: 07/27/20 Analyzed: 07/28/20										
Iron, Total	ND	0.150	mg/L							
Manganese, Total	ND	0.0100	mg/L							
LCS (B0G3557-BS1)										
Prepared: 07/27/20 Analyzed: 07/28/20										
Iron, Total	0.230	0.150	mg/L	0.250		91.9	85-115			
Manganese, Total	0.215	0.0100	mg/L	0.250		86.1	85-115			
LCS (B0G3557-BS2)										
Prepared: 07/27/20 Analyzed: 07/28/20										
Iron, Total	0.229	0.150	mg/L	0.250		91.4	85-115			
Manganese, Total	0.213	0.0100	mg/L	0.250		85.1	85-115			
LCS (B0G3557-BS3)										
Prepared: 07/27/20 Analyzed: 07/28/20										
Iron, Total	0.222	0.150	mg/L	0.250		88.6	85-115			
Manganese, Total	0.266	0.0100	mg/L	0.250		107	85-115			
MRL Check (B0G3557-MRL1)										
Prepared: 07/27/20 Analyzed: 07/28/20										
Iron, Total	0.122		mg/L	0.150		81.5	70-130			
Manganese, Total	0.00922		mg/L	0.0100		92.2	70-130			
MRL Check (B0G3557-MRL2)										
Prepared: 07/27/20 Analyzed: 07/28/20										
Iron, Total	0.120		mg/L	0.150		80.1	70-130			
Manganese, Total	0.00913		mg/L	0.0100		91.3	70-130			
Matrix Spike (B0G3557-MS1)										
Source: C0G3552-01 Prepared: 07/27/20 Analyzed: 07/28/20										
Iron, Total	0.309	0.150	mg/L	0.250	0.0781	92.4	70-130			
Manganese, Total	0.224	0.0100	mg/L	0.250	0.00934	86.0	70-130			
Matrix Spike (B0G3557-MS2)										
Source: C0G3802-01 Prepared: 07/27/20 Analyzed: 07/28/20										
Iron, Total	0.236	0.150	mg/L	0.250	0.00321	93.2	70-130			
Manganese, Total	0.279	0.0100	mg/L	0.250	0.0656	85.2	70-130			

Eastex Environmental Laboratory - Coldspring

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety. Alkalinity titrated to pH 4.5 endpoint.

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved

WQP Report V.3 05162019

Print date and time: 7/29/2020 11:16:40AM

Page 6 of 10



P.O. Box 1089 Coldspring Tx 77331

Website: eastexlabs.com

Email: eastexlab@eastex.net

Tel: 936 653 3249



Eastex Environmental Laboratory - Nacogdoches

1119 South University Drive

Nacogdoches TX, 75961

TCEQ Report ID: **PWS_TX1740014_AC_20200714_WQP Analysis Report**

EPA 200.7 - Quality Control

Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch B0G3557 - EPA 200.7

Matrix Spike (B0G3557-MS3)

Source: C0G4841-01

Prepared: 07/27/20 Analyzed: 07/28/20

Iron, Total	0.249	0.150	mg/L	0.250	0.0289	88.0	70-130			
Manganese, Total	0.218	0.0100	mg/L	0.250	0.0147	81.5	70-130			

Matrix Spike Dup (B0G3557-MSD1)

Source: C0G3552-01

Prepared: 07/27/20 Analyzed: 07/28/20

Iron, Total	0.300	0.150	mg/L	0.250	0.0781	88.9	70-130	2.87	20	
Manganese, Total	0.217	0.0100	mg/L	0.250	0.00934	83.1	70-130	3.27	20	

Matrix Spike Dup (B0G3557-MSD2)

Source: C0G3802-01

Prepared: 07/27/20 Analyzed: 07/28/20

Iron, Total	0.230	0.150	mg/L	0.250	0.00321	90.7	70-130	2.68	20	
Manganese, Total	0.275	0.0100	mg/L	0.250	0.0656	83.6	70-130	1.45	20	

Matrix Spike Dup (B0G3557-MSD3)

Source: C0G4841-01

Prepared: 07/27/20 Analyzed: 07/28/20

Iron, Total	0.257	0.150	mg/L	0.250	0.0289	91.4	70-130	3.39	20	
Manganese, Total	0.225	0.0100	mg/L	0.250	0.0147	84.1	70-130	2.96	20	

Batch B0G3559 - EPA 200.7

Blank (B0G3559-BLK1)

Prepared: 07/27/20 Analyzed: 07/28/20

Calcium, Total	ND	0.500	mg/L							12
Sodium, Total	ND	0.500	mg/L							

Blank (B0G3559-BLK2)

Prepared: 07/27/20 Analyzed: 07/28/20

Calcium, Total	ND	0.500	mg/L							12
Sodium, Total	ND	0.500	mg/L							

Blank (B0G3559-BLK3)

Prepared: 07/27/20 Analyzed: 07/28/20

Calcium, Total	ND	0.500	mg/L							12
Sodium, Total	ND	0.500	mg/L							



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Eastex Environmental Laboratory - Nacogdoches
1119 South University Drive
Nacogdoches TX, 75961

TCEQ Report ID: **PWS_TX1740014_AC_20200714_WQP Analysis Report**

EPA 200.7 - Quality Control

Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B0G3559 - EPA 200.7										
LCS (B0G3559-BS1) Prepared: 07/27/20 Analyzed: 07/28/20										
Calcium, Total	22.9	0.500	mg/L	25.0		91.7	85-115			12
Sodium, Total	24.4	0.500	mg/L	25.0		97.5	85-115			
LCS (B0G3559-BS2) Prepared: 07/27/20 Analyzed: 07/28/20										
Calcium, Total	22.3	0.500	mg/L	25.0		89.3	85-115			12
Sodium, Total	24.1	0.500	mg/L	25.0		96.3	85-115			
LCS (B0G3559-BS3) Prepared: 07/27/20 Analyzed: 07/28/20										
Calcium, Total	22.2	0.500	mg/L	25.0		88.7	85-115			12
Sodium, Total	23.7	0.500	mg/L	25.0		95.0	85-115			
MRL Check (B0G3559-MRL1) Prepared: 07/27/20 Analyzed: 07/28/20										
Calcium, Total	0.453		mg/L	0.500		90.6	70-130			12
Sodium, Total	0.466		mg/L	0.500		93.1	70-130			
MRL Check (B0G3559-MRL2) Prepared: 07/27/20 Analyzed: 07/28/20										
Calcium, Total	0.457		mg/L	0.500		91.4	70-130			12
Sodium, Total	0.468		mg/L	0.500		93.5	70-130			
Matrix Spike (B0G3559-MS1) Source: C0G3552-01 Prepared: 07/27/20 Analyzed: 07/28/20										
Calcium, Total	50.2	2.50	mg/L	25.0	29.0	84.9	70-130			12
Sodium, Total	95.7	2.50	mg/L	25.0	71.6	96.3	70-130			
Matrix Spike (B0G3559-MS2) Source: C0G3802-01 Prepared: 07/27/20 Analyzed: 07/28/20										
Calcium, Total	95.5	2.50	mg/L	25.0	75.1	81.6	70-130			12
Sodium, Total	53.6	2.50	mg/L	25.0	29.1	98.2	70-130			
Matrix Spike (B0G3559-MS3) Source: C0G4841-01 Prepared: 07/27/20 Analyzed: 07/28/20										
Calcium, Total	81.1	2.50	mg/L	25.0	59.1	87.6	70-130			12
Sodium, Total	62.0	2.50	mg/L	25.0	37.7	97.2	70-130			



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Eastex Environmental Laboratory - Nacogdoches
1119 South University Drive
Nacogdoches TX, 75961

TCEQ Report ID: **PWS_TX1740014_AC_20200714_WQP Analysis Report**

EPA 200.7 - Quality Control

Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B0G3559 - EPA 200.7										
Matrix Spike Dup (B0G3559-MSD1)		Source: C0G3552-01		Prepared: 07/27/20 Analyzed: 07/28/20						
Calcium, Total	54.6	2.50	mg/L	25.0	29.0	103	70-130	8.40	20	12
Sodium, Total	103	2.50	mg/L	25.0	71.6	126	70-130	7.36	20	
Matrix Spike Dup (B0G3559-MSD2)		Source: C0G3802-01		Prepared: 07/27/20 Analyzed: 07/28/20						
Calcium, Total	100	2.50	mg/L	25.0	75.1	101	70-130	4.98	20	12
Sodium, Total	55.0	2.50	mg/L	25.0	29.1	104	70-130	2.58	20	
Matrix Spike Dup (B0G3559-MSD3)		Source: C0G4841-01		Prepared: 07/27/20 Analyzed: 07/28/20						
Calcium, Total	83.5	2.50	mg/L	25.0	59.1	97.6	70-130	3.01	20	12
Sodium, Total	65.8	2.50	mg/L	25.0	37.7	112	70-130	5.89	20	



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Eastex Environmental Laboratory - Nacogdoches
1119 South University Drive
Nacogdoches TX, 75961

TCEQ Report ID: **PWS_TX1740014_AC_20200714_WQP Analysis Report**

Qualifiers and Notes

12 CCV associated with sample batch did not meet acceptance criteria.



TEXAS COMMISSION ON ENVIRONMENTAL QUALITY
WATER QUALITY PARAMETER MONITORING FORM 20679

PWS-1740014-AC-30200714

Completed by PWS (or Agent)

Completed by Laboratory

PWS Name:	Refer to Original WQP MF				Laboratory Name:	Eastex Environmental Lab	
PWS ID#:	TX 1740014, Lilly Grove SUD, N060813				ICED Lab ID #:	TX203	
PWS Address:					Laboratory Address:	PO Box 1089	
PWS Contact:					Laboratory Contact:	Cold Spring, TX 77331	
PWS Contact Phone #:					Laboratory Contact Phone #:	Natalia Bondar	
Inhibition of Stabilizer Used (Y/N):	Phosphate	Silica	Calcium carbonate	Laboratory Contact Email:	(936) 653-3249		
TREATMENT:	Aluminum Dosage Rate:	Inhibition Dosage Rate:					
Sample Type (Y):	Compliance	Non-Compliance	Accidental	Lab ID	AL		
Sample Collected (Y):	Public Water System	Accidental	Lab ID				
Temperature and pH (Y/N):	Accidental	Lab ID					
Sample Point ID (e.g. DSU1, BW01)	Sample Location	Date (MM/DD/YY)	Time - 24 hr (HH:MM)	pH	Temp (°C)	Replacement?	
PWS005 EWAP	North Plant	07/14/20	1040				
PWS002 EWAP	Nat Plant	07/14/20	1025				
DS01 DSTWAP	366 CR 820	07/14/20	1110				
DS01 DSTWAP	12232 FM 343	07/14/20	1035				
Laboratory Comments:							
Parameters Requested: Analyses are reduced for the parameters checked. * If Inhibition of Stabilizer or Silica are used, then these parameters should also be checked depending on which is in use.							
Lab Sample ID							
005-1538-01							
005-1538-02							
005-1538-03							
005-1538-04							
005-1538-05							
005-1538-06							
005-1538-07							
005-1538-08							
005-1538-09							
005-1538-10							
005-1538-11							
005-1538-12							
005-1538-13							
005-1538-14							
005-1538-15							
005-1538-16							
005-1538-17							
005-1538-18							
005-1538-19							
005-1538-20							
005-1538-21							
005-1538-22							
005-1538-23							
005-1538-24							
005-1538-25							
005-1538-26							
005-1538-27							
005-1538-28							
005-1538-29							
005-1538-30							
005-1538-31							
005-1538-32							
005-1538-33							
005-1538-34							
005-1538-35							
005-1538-36							
005-1538-37							
005-1538-38							
005-1538-39							
005-1538-40							
005-1538-41							
005-1538-42							
005-1538-43							
005-1538-44							
005-1538-45							
005-1538-46							
005-1538-47							
005-1538-48							
005-1538-49							
005-1538-50							
005-1538-51							
005-1538-52							
005-1538-53							
005-1538-54							
005-1538-55							
005-1538-56							
005-1538-57							
005-1538-58							
005-1538-59							
005-1538-60							
005-1538-61							
005-1538-62							
005-1538-63							
005-1538-64							
005-1538-65							
005-1538-66							
005-1538-67							
005-1538-68							
005-1538-69							
005-1538-70							
005-1538-71							
005-1538-72							
005-1538-73							
005-1538-74							
005-1538-75							
005-1538-76							
005-1538-77							
005-1538-78							
005-1538-79							
005-1538-80							
005-1538-81							
005-1538-82							
005-1538-83							
005-1538-84							
005-1538-85							
005-1538-86							
005-1538-87							
005-1538-88							
005-1538-89							
005-1538-90							
005-1538-91							
005-1538-92							
005-1538-93							
005-1538-94							
005-1538-95							
005-1538-96							
005-1538-97							
005-1538-98							
005-1538-99							
005-1538-100							
005-1538-101							
005-1538-102							
005-1538-103							
005-1538-104							
005-1538-105							
005-1538-106							
005-1538-107							
005-1538-108							
005-1538-109							
005-1538-110							
005-1538-111							
005-1538-112							
005-1538-113							
005-1538-114							
005-1538-115							
005-1538-116							
005-1538-117							
005-1538-118							
005-1538-119							
005-1538-120							
005-1538-121							
005-1538-122							
005-1538-123							
005-1538-124							
005-1538-125							
005-1538-126							
005-1538-127							
005-1538-128							
005-1538-129							
005-1538-130							
005-1538-131							
005-1538-132							
005-1538-133							
005-1538-134							
005-1538-135							
005-1538-136							
005-1538-137							
005-1538-138							
005-1538-139							
005-1538-140							
005-1538-141							
005-1538-142							
005-1538-143							
005-1538-144							
005-1538-145							
005-1538-146							
005-1538-147							
005-1538-148							
005-1538-149							
005-1538-150							
005-1538-151							
005-1538-152							
005-1538-153							
005-1538-154							
005-1538-155							
005-1538-156							
005-1538-157							
005-1538-158							
005-1538-159							
005-1538-160							
005-1538-161							
005-1538-162							
005-1538-163							
005-1538-164							
005-1538-165							
005-1538-166							
005-1538-167							
005-1538-168							
005-1538-169							
005-1538-170							
005-1538-171							
005-1538-172							
005-1538-173							
005-1538-174							
005-1538-175							
005-1538-176							
005-1538-177							
005-1538-178							
005-1538-179							
005-1538-180							
005-1538-181							
005-1538-182							
005-1538-183							
005-1538-184							
005-1538-185							
005-1538-186							
005-1538-187							
005-1538-188							
005-1538-189							
005-1538-190							
005-1538-191							
005-1538-192							
005-1538-193							
005-1538-194							
005-1538-195							
005-1538-196							
005-1538-197							
005-1538-198							
005-1538-199							
005-1538-200							
005-1538-201							
005-1538-202							
005-1538-203							
005-1538-204							
005-1538-205							
005-1538-206							
005-1538-207							
005-1538-208							
005-1538-209							
005-1538-210							
005-1538-211							
005-1538-212							
005-1538-213							
005-1538-214							
005-1538-215							
005-1538-216							
005-1538-217							
005-1538-218							
005-1538-219							
005-1538-220							
005-1538-221							
005-1538-222							
005-1538-223							
005-1538-224							
005-1538-225							
005-1538-226							
005-1538-227							
005-1538-228							
005-1538-229							
005-1538-230							
005-1538-231							
005-1538-232							
005-1538-233							
005-1538-234							
005-1538-235							
005-1538-236							
005-1538-237							
005-1538-238							
005-1538-239							
005-1538-240							
005-1538-241							
005-1538-242							
005-1538-243							
005-1538-244							
005-1538-245							
005-1538-246							
005-1538-247							
005-1538-248							
005-1538-249							
005-1538-250							
005-1538-251							
005-1538-252							
005-1538-253							
005-1538-254							
005-1538-255							
005-1538-256							
005-1538-257							
005-1538-258							
005-1538-259							
005-1538-260							
005-1538-261							
005-1538-262							
005-1538-263							
005-1538-264							
005-1538-265							
005-1538-266							
005-1538-267							
005-1538-268							
005-1538-269							
005-1538-270							
005-1538-271							
005-1538-272							
005-1538-273							
005-1538-274							
005-1538-275							
005-1538-276							
005-1538-277							
005-1538-278							
005-1538-279							
005-1538-280							
005-1538-281							
005-1538-282							
005-1538-283							
005-1538-284							
005-1538-285							
005-1538-286							
005-1538-287							
005-1538-288							
005-1538-289							
005-1538-290							
005-1538-291							
005-1538-292							
005-1538-293							
005-1538-294							
005-1538-295							
005-1538-296							
005-1538-297							
005-1538-298							
005-1538-299							
005-1538-300							
005-1538-301							
005-1538-302							
005-1538-303							
005-1538-304							
005-1538-305							
005-1538-306							
005-1538-307							
005-1538-308							
005-1538-309							
005-1538-310							
005-1538-311							
005-1538-312							
005-1538-313							
005-1538-314							
005-1538-315							
005-1538-316							
005-1538-317							
005-1538-318							
005-1538-319							
005-1538-320							
005-1538-321							
005-1538-322							
005-1538-323							
005-1538-324							
005-1538-325							
005-1538-326							
005-1538-327							
005-1538-328							
005-1538-329							
005-1538-330							
005-1538-331							
005-1538-332							
005-1538-333							
005-1538-334							
005-1538-335							
005-1538-336							
005-1538-337							
005-1538-338							
005-1538-339							
005-1538-340							
005-1538-341							
005-1538-342							
005-1538-343							
005-1538-344							
005-1538-345							
005-1538-346							
005-1538-347							
005-1538-348							
005-1538-349							
005-1538-350							
005-1538-351							
005-1538-352							
005-1538-353							
005-1538-354							
005-1538-355							
005-1538-356							
005-1538-357							
005-1538-358							
005-1538-359							
005-1538-360							
005-1538-361							
005-1538-362							
005-1538-363							
005-1538-364							
005-1538-365							
005-1538-366							
005-1538-367							
005-1538-368							
005-1538-369							
005-1538-370							
005-1538-371							
005-1538-372							
005-1538-373							
005-1538-374							
005-1538-375							
005-1538-376							
005-1538-377							
005-1538-378							
005-1538-379							
005-1538-380							
005-1538-381							
005-1538-382							
005-1538-383							
005-1538-384							
005-1538-385							
005-1538-386							
005-1538-387							
005-1538-388							
005-1538-389							
005-1538-390							
005-1538-391							
005-1538-392							
005-1538-393							
005-1538-394							
005-1538-395							
005-1538-396							
005-1538-397							
005-1538-398							
005-1538-399							
005-1538-400							
005-1538-401							
005-1538-402							
005-1538-403							
005-1538-404							
005-1538-405							
005-1538-406							
005-1538-407							
005-1538-408							
005-1538-409							
005-1538-410							
005-1538-411							
005-1538-412							
005-1538-413							
005-1538-414							
005-1538-415							
005-1538-416							
005-1538-417							
005-1538-418							
005-1538-419							
005-1538-420							
005-1538-421							
005-1538-422							
005-1538-423							
005-1538-424							
005-1538-425							
005-1538-426							
005-1538-427							
005-1538-428							
005-1538-429							
005-1538-430							
005-1538-431							
005-1538-432							
005-1538-433							
005-1538-434							
005-1538-435							
005-1538-436							
005-1538-437							
005-1538-438							
005-1538-439							
005-1538-440							
005-1538-441							
005-1538-442							
005-1538-443							
005-1538-444							
005-1538-445							
005-1538-446							
005-1538-447							
005-1538-448							
005-1538-449							
005-1538-450							
005-1538-451							
005-1538-452							
005-1538-453							
005-1538-454							
005-1538-455							
005-1538-456							
005-1538-457							
005-1538-458							
005-1538-459							
005-1538-460							
005-1538-461							
005-1538-462							
005-1538-463							
005-1538-464							
005-1538-465							
005-1538-466							
005-1538-467							
005-1538-468							
005-1538-469							
005-1538-470							
005-1538-471							
005-1538-472							
005							



TEXAS COMMISSION ON ENVIRONMENTAL QUALITY
WATER QUALITY PARAMETER MONITORING FORM 20679

Completed by Laboratory

Completed by PWS (or Agent)

COPY

PWS Name: Lilly Grove SCD
PWS ID #: TX 1740014
PWS Address: 7435 FM 1638, Nacogdoches, TX 75988
PWS Contact: Boyd Deebay
PWS Contact Phone #: 936-569-9292
Inhibitor or Stabilizer Used (Y/N): ☒ Phosphate
TREATMENT: Alkalinity Dosage Rate: 1
Inhibitor Dosage Rate: 1
Laboratory Name: Eastern Environmental Lab
TCEQ Lab ID #: TX 009162
Laboratory Address: PO Box 63375
Nacogdoches TX 75961
Laboratory Contact: Paul Hughes
Laboratory Contact Phone #: 936-569-8879

Sample Type (Y/N):	☒ Compliance	Non-compliance
Sample Collector (Y/N):	☒ Public Water System	Accredited Lab
Temperature and pH (Y or N):	☒ Are temperature and pH included on the sampling entity's Laboratory Approval Form on file at the TCEQ?	3rd Party Contractor --> LAB ID AL

Sample Information	Where temperature and pH measured in the field within 15 minutes of sample collection?						
Sample Collection	Field Measurements						
Sample Point ID (e.g., D501, PBCU001)	Date (MMDDYY)	Time - 24 hr (HHMM)	pH	Temp (°C)	Replacement?	Original Sample ID #	Original Sample Date (MMDDYY)

PSCU005 ENDR	Martin Plant	7/14/20	10:40A	8.1	17.8°C		
PSCU002 ENDR	Nat Plant	7/14/20	10:25A	8.0	22.2°C		
D501	D51WBQ	366 CR 820	7/14/20	11:10A	7.6	19.5°C	
D501	D51WBQ	12232 FH 343	7/14/20	10:35A	7.6	21.1°C	

Lab Sample ID	Alkalinity (1927)	Calcium (1919)	Chloride (1017)	Conductivity (1064)	Hardness (1915)	Iron (1028)	Manganese (1032)	Sodium (1052)	Sulfate (1055)	TDS (1930)	O-Phosphate (1044) *	Silica (1049) *
NOG0813-01	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓
NOG0813-02	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓
NOG0813-03	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓
NOG0813-04	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓

I acknowledge that information on this form is true and correct and sites selected for sampling follow the instructions in the TCEQ Monitoring and Sample Collection Guidance for Water Quality Parameters. This includes, but not limited to the measurement of pH and temperature immediately upon collection. Falsification of this form or tampering with water samples is a crime punishable under state and/or federal law. (Texas Penal Code, Title 8, Chapter 37.10)

Sample Conditions Upon Receipt (Y)

Name of Authorized PWS Representative (Print)	Signature	Organization	Date
Boyd Deebay	Boyd Deebay	Lilly Grove	7/15/2020

Rejection Code (if applicable):	Actual / Corrected sample temperature:	Thermometer ID #:
	9.8 / 9.8	017

Relinquished By (Signature)	Date/Time:	Relinquished By Counter (Signature)	Date/Time:
Boyd Deebay	7/15/20 1040	Adrian Hunter	07/15/20 1040

Laboratory Comments:



EASTEX ENVIRONMENTAL LABORATORY, INC.

P.O. Box 1089 • Coldspring, TX 77331
(936) 653-3249 • (800) 525-0508
P.O. Box 631375 • Nacogdoches, TX 75963-1375
(936) 569-8879 • FAX (936) 569-8951
www.eastexlabs.com

White Copy-Follows Samples
Yellow Copy-Laboratory
Pink Copy-Client Copy

REPORT TO:

INVOICE TO:

Company: Lilly Grove Sued

Company:

Address:

Address:

Attn:

Attn:

Phone#:

Phone#:

Email:

INSTRUCTIONS:

P.O. #:

C or G: C = Composite G = Grab

Sampler's Name (print):

Matrix: DW=Drinking Water WW=Wastewater SO=Soil/Sludge OT=Other

Sampler's Signature:

Container Size: 1=Gallon 2=1/2 Gallon 3=Quart/Liter 4=500mL 5=250mL
6=125mL (4oz) 7=60mL (2 oz) 8=40mL Vial 9=Other

Project Name:

Type: P = Plastic G = Glass T = Teflon S = Sterile
Preservatives: C=Chilled S=Sulfuric Acid N=Nitric Acid B=Base/Caustic Z=Zn Acetate
ST=Sodium Thiosulfate H=HCL O=Other

Work Order ID

Sample ID

Date

Time

Matrix

C or G

DO

pH

Cl2

Flow

Temp

#

Size

Type

Pres

Field Data

Containers

ANALYSIS REQUESTED

Remarks: No GOSB-01,
COCY533-01 Matrix
Plant was not
Subcontracted over
Justin re-submitted
work order all together
on 7/22/20.

WQP

Relinquished By:

Received By:

Date

Time

Relinquished By:

Received By:

Date

Time

Relinquished By:

Received By and/or Checked in By:

Date

Time

LAB USE ONLY

Sample Condition Acceptable:

YES / NO

Temp °C

*Therm ID

Logged in By:

Time

Alternate Check In:

Date

Time

Temp °C

*Therm ID

Logged in By:

Time

Received Iced: YES / NO

Received Iced: YES / NO

Received Iced: YES / NO

*Thermometer has 0.0 factor and recorded temperature is actual temperature